

CERTIFICATION OF CALIBRATION



Date Of Calibration: 14-Feb-2020

Certificate Number: G505522_2/25062

Issued by: QED Environmental Systems Ltd.

Customer: Enitial Projects Limited
Enterprise Drive Four Ashes Wolverhampton
Staffordshire WV10 7DE UNITED KINGDOM

Description: Gas Analyser

Model: GA5000

Serial Number: G505522

UKAS Accredited results:

Results after adjustment :

Methane (CH ₄)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
5.0	4.8	0.41
15.0	14.9	0.64
50.0	49.6	0.94

Carbon Dioxide (CO ₂)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
5.0	4.8	0.43
15.0	14.7	0.70
50.0	50.3	1.1

Oxygen (O ₂)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
21.3	21.4	0.31

The inwards assessment was carried out 12-Feb-2020.

The maximum adjustment is larger than the inwards assessment uncertainty.

Inwards assessment data is available if requested.

All concentrations are molar.

CH₄, CO₂ readings recorded at : 31.2 °C ± 2.5 °C

O₂ readings recorded at : 22.7 °C ± 2.5 °C

Barometric Pressure : 1005 mbar ± 4 mbar

Method of Test : The analyser is calibrated in a temperature controlled chamber using a series of reference gases, in compliance with procedure LP004.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Instance:108 IGC Instance:107

Page 1 of 2 | LP015GIUKAS-2.4

www.qedenv.com +44 (0) 333 800 0088 sales@qedenv.co.uk

QED Environmental Systems Ltd. Cyan Park - Unit 3, Jimmy Hill Way, Coventry, CV2 4QP, UNITED KINGDOM

Registered in England and Wales 1898734

SERVICE REPORT



Issued by: QED Environmental Systems Ltd.

Customer Name: Enital Projects Limited

Model: GA5000

Part Number: GA5K0AC0-101

Serial Number: G505522

Date of Service: 17-Feb-2020

Service Engineer: David Edwards

Calibration Engineer: Suk Balrey

Verification / Approved By: Dawn Hemings

Signature:

Service Comments/Feedback:

Thank you for returning your gas analyser to the Geotech Service Centre for full service and calibration. We are pleased to inform you that the analyser was received and fully assessed by one of our experienced Service Engineers and no significant faults or issues were observed during the assessment.

Our extensive service was carried out, and any necessary components replaced.

We are pleased to inform you that we have installed the latest version of firmware on your gas analyser, so you will be able to benefit from the latest features and improvements.

The analyser has successfully passed all of our rigorous testing and quality checks and has been calibrated using our bespoke, state of the art calibration facility.

For further information about how to get the best use from your instrument please visit our YouTube channel <http://www.youtube.com/GeotechTV> and on our Website <http://www.geotechuk.com>

If you require any further assistance with your instrument please email our Technical Support Team at technical@geotech.co.uk or call us on +44 1926 338111 (Monday to Thursday 08.30 - 17.00 & Friday 08.30 - 15.30) UK time zone.

Next Service Due: 14-Aug-2020

17025

Our ISO accreditation for our customised auto-calibration facilities

5

Number of days we aim to complete your service within

50

Number of checks instruments are subject to when serviced

65

Number of countries from which we service instruments /accessories each year

7,384

Number of calibrations completed in last 12 months

340

Minimum number of service instruments we process each month

25

Cost (in £) of fully insured analyser collection for our UK customers

www.qedenv.com

+44 (0) 333 800 0088

sales@qedenv.co.uk

QED Environmental Systems Ltd. Cyan Park - Unit 3, Jimmy Hill Way, Coventry, CV2 4QP, UNITED KINGDOM

Registered in England and Wales 1898734