Cirrus Research plc

Measurement Summary Report

Name	4. Investment Managment MGP
Time	07/09/2015 13:28:40
Duration	00:00:20
Instrument	G066183, CR:171B

Summary

LAeq	60.4 dB
LCPeak	96.7 dB
C-A	10.8 dB
LEPd	28.8 dB
LAFMax	66.2 dB

Projected Exposure

30 Mins	48.4 dB
1 Hour	51.4 dB
2 Hours	54.4 dB
3 Hours	56.2 dB
4 Hours	57.4 dB


Projected Exposure

5 Hours	58.4 dB
6 Hours	59.2 dB
7 Hours	59.8 dB
8 Hours	60.4 dB
10 Hours	61.4 dB
12 Hours	62.2 dB


Calibration Information

07/09/2015 13:19:22 -0.34 dB

Time History



Frequency Bands



Report ID