

TEST DATE AND CONDITIONS			
Date		25/06/2024	
Atmospheric Pressure		999	mB
Ambient Temperature		21.9	°C
EnviroNics Serial No.		5089	

**GFM436 Final Inspection & Calibration
Check Certificate**

GAS DATA LTD	
Unit D	
Earlplace Business Park	
Fletchamstead Highway	
Coventry	
CV4 9XL	
Tel 02476303311	Fax 02476307711



GAS DATA
LISTEN • ANALYSE • INFORM

Customer	Curtins Consulting
Certificate Number	125304
Order Number	338558

Serial Number	13712
Software Version	G436-00.0029/0010

Recalibration DUE Date
25/06/25

Instrument Checks					
Keyboard	✓		Display Contrast	✓	
Pump Flow In	550	Accept > 200 cc/min	Pump Flow @ -200mB	350	Accept > 200 cc/min
Clock Set / Running	✓		Labels Fitted	✓	

Gas Checks						
Sensor	CH ₄		CO ₂		O ₂	
	Instrument Gas	True Gas Value %	Instrument Gas	True Gas Value %	Instrument Gas	True Gas Value %
	Readings %		Readings %		Readings %	
	59.7	60	40.2	40	20.9	20.9
	Accept ±3.0		Accept ±3.0		Accept ±0.5	
	5.0	5	4.9	5	6.0	6
Zero Reading 100% N2	0.0	0	0.0	0	0.0	0
	Accept ±0.0		Accept ±0.0		Accept ±0.1	

Optional Gas Checks						
Applied Gas & Range		Concentration Tested @ (ppm)	Instrument Readings (ppm)			
Gas Type	Range (ppm)		Zero Reading		Instrument Gas Reading	
H2S	5000	1500	0	Accept ±0.0	1500	Accept ±5.0%
CO	2000	1000	0	Accept ±0.0	997	Accept ±5.0%
Hexane	2.0%	2.0%	0	Accept ±0.0	1.99	Accept ±10.0%

Cross Gas Effects									
Applied Gas (ppm)		Instrument Readings (ppm)							
Gas Type	Concentration	Toxic 1:	H2S	Toxic 2:	CO	Toxic 3:	HEX		
H2S	1500	1500		0		0			
CO	1000	100		997		0			
Hexane	2.0%	0		0		1.99			

Pressure Checks			
Atmospheric Pressure [AP] (mB)			
Current Atmospheric Pressure (mB)		Instrument Atmospheric Pressure Reading (mB)	
AP Open Ports		999	Accept ± 2.0
AP Port (Internal)	+800 mB	800	Accept ± 5.0
	+1200mB	1200	Accept ± 5.0

Flow Checks					
Borehole Flow			Differential Pressure		
Applied Reading (l/h)	Instrument Reading (l/h)		Applied Pressure (Pa)	Instrument Reading (Pa)	
-30	-29.9	Accept ± 3.0	-299	-298	Accept ± 50
-3	-3.0	Accept ± 1.0	-14	-14	Accept ± 6.0
0	0.0	Accept ± 0.0	0	0	Accept ± 0.5
3	3.0	Accept ± 0.5	17	16	Accept ± 3.0
30	30.6	Accept ± 3.0	407	420	Accept ± 50
60	59.8	Accept ± 6.0	1275	1265	Accept ± 130
90	91.9	Accept ± 9.0	2550	2652	Accept ± 250

Temperature Checks		
Calibration Temperature	Instrument Temperature Reading $^{\circ}\text{C}$	
Applied Temperature $^{\circ}\text{C}$		
-10	-10.0	Accept ± 2.0
0	0.0	Accept ± 1.0
30	30.0	Accept ± 1.0
60	60.0	Accept ± 1.0
100	100.0	Accept ± 1.0

Technician:

Date Tested:
25/06/2024

The instrument identified by the serial number stated above has been tested by Gas Data personnel for calibration accuracy on the date and under the ambient conditions stated. Gas Data Ltd internal BS EN ISO9001:2015, BS EN ISO14001:2015, BS EN ISO45001:2018 compliant workshop procedures were followed to apply known calibration test gases, gas flow rates, pressures and temperatures of the values stated. The results displayed on the instrument at each stage are recorded above.


TEST DATE AND CONDITIONS			
Date	25.6.24		
Atmospheric Pressure	999	mB	
Ambient Temperature	21.9	°C	
EnviroNics Serial No.	5089		

GAS DATA LTD Unit D Earlplace Business Park Fletchamstead Highway Coventry CV4 9XL UK +44 (0) 24 7630 3311	 GAS DATA LISTEN · ANALYSE · INFORM
---	--

GFM436-1 OUTWARD INSPECTION & QUALITY CHECK SHEET

INSTRUMENT DETAILS			
SO Number	Instrument Type	Instrument Serial Number + SW Version	Job Number(s)
338558	GFM436	13712 G436-27/10	125304

Calibration Technician



Date ...25.6.24...

Inspection Technician

Date ...26.6.24...

INSTRUMENT CHECKS		Pass (P), Fail (F) or not applicable (NA)	INSTRUMENT PACKING LIST		Tick if included
Function Tests	Dust Caps Fitted	P	Instrument		✓
	Keyboard Test (All Keys)	P	Leather Case		X
	Backlight	P	Instrument Strap		✓
	Clock Set / Running	P	AC Battery Charger (UK)		✓
	Comms Test	P	AC Battery Charger (EURO)		X
	Pump Flow Test (In & Out)	P	AC Battery Charger (US)		X
	Overall Leak Test (30mB)	n/a	AC Battery Charger (AUS)		X
	Battery Charge Test	P	Gas Sample Pipe - (new issue)		✓
	Service Date set to?	25.6.25	Flow Sample Pipe - (new issue)		✓
Channel Tests	Data Logging Enabled?	P	Hard Carry Case		✓
	Verify CH4/LEL/Hexane/PID	P	Spares Pot		✓
	Verify CO2	P	Allen Key		X
	Verify O2	P	Temperature Probe		X
	Verify H2S	P	Vane Anemometer		X
	Verify CO	P	USB Cable		✓
	Verify LEL	P	USB Memory stick		X
	Verify 1 st Option Gas	N/A	SM V5 Software	Ver 6.05	X
	Verify Atmospheric pressure	P	Internal Filter Pack	Qty	X
	Verify differential pressure	P	External Filter Pack	Qty	X
	Verify flow	P	Field Guide		X
	Verify temperature probe input	P	Extra Items: Sample Tube		
	Verify vane anemometer input	P			
	DataBase Checks	Jobcard(s) completed and signed	P	Comments:	
Jobcard(s) booked off database		P			
Calibration certificate completed		✓			
Complete & print QI record		n/a			
Label Checks	No. of Calibration label fitted	GDC 13422			
	MCERTS label displayed	N/A			
	Warranty label fitted				
H2S Range	H2S Range from Sales Order	5000 ppm			
	H2S Range from Cal Cert	5000 ppm			
	Over-range value correct?	P			
QA Cylinders Used	No. of High Range CH4/CO2	Silcyl 110 - 57522			
	No. of Low Range CH4/CO2/O2	Silcyl 110 - 57550			
	No. of H2S	Silcyl 110 - 6032502			
	No. of CO	Silcyl 110 - 58428			
	No. of Optional Cylinder	N/A			

