

TEST DATE AND CONDITIONS			
Date		14/08/19	
Atmospheric Pressure		993	mB
Ambient Temperature		21.9	°C
EnviroNics Serial No.		5089	

**GFM436 Final Inspection & Calibration
Check Certificate**

GAS DATA LTD

Unit 4, Fairfield Court

Seven Stars Estate

Wheler Rd

Coventry

CV3 4LJ

Tel 02476303311

Fax 02476307711



Customer	WML Consulting Ltd
Certificate Number	121220
Order Number	323564

Serial Number	12969
Software Version	G436-00.0027/0010

Recalibration DUE Date
14/08/20

Instrument Checks					
Keyboard	✓		Display Contrast	✓	
Pump Flow In	500	Accept > 200 cc/min	Pump Flow @ -200mB	300	Accept > 200 cc/min
Clock Set / Running	✓		Labels Fitted	✓	

Gas Checks						
Sensor	CH ₄		CO ₂		O ₂	
	Instrument Gas	True Gas Value %	Instrument Gas	True Gas Value %	Instrument Gas	True Gas Value %
	Readings %		Readings %		Readings %	
	60	60	40	40	20.9	20.9
	Accept ±3.0		Accept ±3.0		Accept ±0.5	
Zero Reading 100% N2	5	5	5	5	6	6
	Accept ±0.3		Accept ±0.3		Accept ±0.3	
Zero Reading 100% N2	0	0	0	0	0	0
	Accept ±0.0		Accept ±0.0		Accept ±0.1	

Optional Gas Checks						
Applied Gas & Range		Concentration Tested @ (ppm)	Instrument Readings (ppm)			
Gas Type	Range (ppm)		Zero Reading		Instrument Gas Reading	
H2S	5000	1500	0	Accept ±0.0	1500	Accept ±5.0
CO	2000	1000	0	Accept ±0.0	1000	Accept ±5.0
Hexane	2.0%	2.0%	0	Accept ±0.0	1.99	Accept ±10.0

Cross Gas Effects								
Applied Gas (ppm)		Instrument Readings (ppm)						
Gas Type	Concentration	Toxic 1:	H2S	Toxic 2:	CO	Toxic 3:	HEX	
H2S	1500	1500		0		0		
CO	1000	40		1000		0		
Hexane	2.0%	0		0		1.99		

Pressure Checks			
Atmospheric Pressure [AP] (mB)			
Current Atmospheric Pressure (mB)		Instrument Atmospheric Pressure Reading (mB)	
AP Open Ports		993	Accept ± 2.0
AP Port (Internal)	+800 mB	800	Accept ± 5.0
	+1200 mB	1200	Accept ± 5.0

Flow Checks					
Borehole Flow			Differential Pressure		
Applied Reading (l/h)	Instrument Reading (l/h)		Applied Pressure (Pa)	Instrument Reading (Pa)	
-30	-30.3	Accept ± 3.0	-556	-570	Accept ± 50
-3	-3	Accept ± 1.0	-23	-22	Accept ± 6.0
0	0	Accept ± 0.0	0	0	Accept ± 0.5
3	3	Accept ± 0.5	17	17	Accept ± 3.0
30	30	Accept ± 3.0	385	384	Accept ± 50
60	60	Accept ± 6.0	1192	1193	Accept ± 130
90	90.4	Accept ± 9.0	2362	>>>>	Accept ± 250

Temperature Checks		
Calibration Temperature	Instrument Temperature Reading °C	
Applied Temperature °C		
-10	-10	Accept ± 2.0
0	0	Accept ± 1.0
30	30	Accept ± 1.0
60	60	Accept ± 1.0
100	100	Accept ± 1.0

Technician:
Jack Rutland

Date Tested:
15/08/19

The instrument identified by the serial number stated above has been tested by Gas Data personnel for calibration accuracy on the date and under the ambient conditions stated. Gas Data Ltd internal BS EN ISO9001:2015 compliant workshop procedures were followed to apply known calibration test gases, gas flow rates, pressures and temperatures of the values stated. The results displayed on the instrument at each stage are recorded above.