CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE No.: 02241451

CLIENT: Green Piling

INSTRUMENT TYPE / No.: V901 / 1451

CALIBRATION DATE: 09/02/2024

CALIBRATION DUE: February 2025

CALIBRATED BY: RJA

CALIBRATION ACCURACY:- @40Hz 5mm/s

	A channel	B channel	VDV channel
Peak Particle Velocity L	±5 %	\ %	X \ %
Peak Particle Velocity V	±5 %	\ %	Y \ %
Peak Particle Velocity T	±5 %	\ %	Z \ %

AIR OVERPRESSURE CHANNEL - Peak Level Unweighted \ dB(Lin)

WE HEREBY CERTIFY THAT THIS SEISMOGRAPH FULLY COMPLIES WITH THE MANUFACTURERS SPECIFICATION

CERTIFIED BY:

DATE: 09/02/2024

THIS CERTIFICATE IS VALID FOR 12 MONTHS

The above calibration was carried out using equipment calibrated as follows:-
Pulsar Acoustic Calibrator 100B, serial number 039467, calibrated November 2023
ISO-TECH IFG 100 Oscillator, serial number 300351, calibrated September 2023
Monitran Vibration Meter, serial number 213608, calibrated September 2023
Precision Gold PG012 Multimeter, serial number 09000182, calibrated September 2023

THIS CALIBRATION IS TRACEABLE TO NATIONAL STANDARDS

VIBROCK LIMITED
Shanakiel
Ilkeston Road
Heanor
Derbyshire DE75 7DR
Tel: 01773 711211
Fax: 01773 711311
Email: vibrock@vibrock.com
Web: www.vibrock.com

VIBROCK

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE No.: 02241029

CLIENT: Green Piling

INSTRUMENT TYPE / No.: V9D1 / 1029

CALIBRATION DATE: 09/02/2024

CALIBRATION DUE: February 2025

CALIBRATED BY: PJA

CALIBRATION ACCURACY:- @40Hz 5mm/s

	A channel	B channel	VDV channel
Peak Particle Velocity L	±5 %	\ %	X \ %
Peak Particle Velocity V	±5 %	\ %	Y \ %
Peak Particle Velocity T	±5 %	\ %	Z \ %

AIR OVERPRESSURE CHANNEL - Peak Level Unweighted \ dB(Lin)

WE HEREBY CERTIFY THAT THIS SEISMOGRAPH FULLY COMPLIES WITH THE MANUFACTURERS SPECIFICATION

CERTIFIED BY:

DATE: 09/02/2024

THIS CERTIFICATE IS VALID FOR 12 MONTHS

The above calibration was carried out using equipment calibrated as follows:-
Pulsar Acoustic Calibrator 100B, serial number 039467, calibrated November 2023
ISO-TECH IFG 100 Oscillator, serial number 300351, calibrated September 2023
Monitran Vibration Meter, serial number 213608, calibrated September 2023
Precision Gold PG012 Multimeter, serial number 09000182, calibrated September 2023

THIS CALIBRATION IS TRACEABLE TO NATIONAL STANDARDS

VIBROCK LIMITED
Shanakiel
Ilkeston Road
Heanor
Derbyshire DE75 7DR
Tel: 01773 711211
Fax: 01773 711311
Email: vibrock@vibrock.com
Web: www.vibrock.com

VIBROCK