

CERTIFICATION OF CALIBRATION



Date Of Calibration: 28-Jun-2022

Certificate Number: G507783_1/30716

Issued by: QED Environmental Systems Ltd.

Customer: L K Consult Limited
Unit 29 Eton Business Park Eton Hill Road Radcliffe
Manchester M26 2ZS UNITED KINGDOM

Description: Gas Analyser

Model: GA5000

Serial Number: G507783

UKAS Accredited results:

Results after adjustment :

Methane (CH ₄)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
5.0	4.9	0.072
15.0	15.0	0.13
60.0	59.7	0.42

Carbon Dioxide (CO ₂)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
5.0	4.9	0.074
15.0	14.8	0.13
40.0	40.0	0.29

Oxygen (O ₂)		
Certified Gas (%)	Instrument Reading (%)	Uncertainty (%)
21.1	21.2	0.25

The inwards assessment was carried out 08-Jun-2022.

The maximum adjustment is larger than the specification limit.

Inwards assessment data is available if requested.

All concentrations are molar.

CH₄, CO₂ readings recorded at : 31.9 °C ± 2.5 °C

O₂ readings recorded at : 23.3 °C ± 2.5 °C

Barometric Pressure : 1005 mbar ± 4 mbar

Method of Test : The analyser is calibrated in a temperature controlled chamber using a series of reference gases, in compliance with procedure LP004.

Instrument has passed calibration as the measurement result is within the specification limit. The specification limit takes into account the measurement uncertainty.

The results relate only to the item calibrated

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Instance:114 IGC Instance:114

Page 1 of 2 | LP015GIUKAS-2.5

www.qedenv.com +44 (0) 333 800 0088 sales@qedenv.co.uk

QED Environmental Systems Ltd. Cyan Park - Unit 3, Jimmy Hill Way, Coventry, CV2 4QP, UNITED KINGDOM

Registered in England and Wales 1898734

CERTIFICATION OF CALIBRATION



Date Of Calibration: 28-Jun-2022

Certificate Number: G507783_1/30716

Issued by: QED Environmental Systems Ltd.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor of $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Calibrations marked 'Non-UKAS Accredited results' on this certificate have been included for completeness.

Non-UKAS accredited results after adjustment:

Barometer (mbar)	
Reference	Instrument Reading
1005	1006

Additional Gas Cells		
Gas	Certified Gas (ppm)	Instrument Reading (ppm)
CO	499	499
H ₂ S	52.6	53

Date of Issue : 29-Jun-2022

Approved by Signatory

Laura McBride

Laboratory Inspection

End of Certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Instance:114 IGC Instance:114

Page 2 of 2 | LP015GIUKAS-2.5

www.qedenv.com +44 (0) 333 800 0088 sales@qedenv.co.uk

QED Environmental Systems Ltd. Cyan Park - Unit 3, Jimmy Hill Way, Coventry, CV2 4QP, UNITED KINGDOM

Registered in England and Wales 1898734

Instrument Service Report

Unit Type: GA5000	Part Number: GA5KB0B0-301	Date: 29-Jun-2022	Next Service Due: 28-Dec-2022	Customer Name: L K Consult Limited
Serial Number: G507783				
Actions/Investigation Description	Result	Comments		
Serial Number Check	Yes			
Battery Requires Replacing	No			
Internal Flow Calibrated	Yes			
Full Automatic Calibration	Pass			
Serial Comms Test (USB)	Pass			
Service history of instrument reviewed	Yes			
Inwards gas check data reviewed	Yes			
Instrument turns on	Pass			
Backlight operates correctly	Yes			
External visual inspection performed	Pass			
Internal visual inspection performed	Pass			
All screws tightened to correct torque	Yes			
All connectors are secure	Pass			
Check diagnostic channels	Pass			
Case compression test	Pass			
Impact and stability test	Pass			
Pressure transducer test(s) as per user operation	Pass			

Instrument Service Report

Unit Type: GA5000	Part Number: GA5KB0B0-301	Date: 29-Jun-2022	Next Service Due: 28-Dec-2022	Customer Name: L K Consult Limited
Serial Number: G507783				
Actions / Investigation Description		Result	Comments	
Final visual inspection on instrument		Pass		
Case assembly closed and screws tightened to correct torque		Yes		
PTFE filters replaced		Yes		
Pump flow greater than 550 ml/min		Pass		
Automated instrument pressure system test (leak test)		Pass		
Pump vacuum greater than -400 mb and flow fails		Pass		
GPS function test		Pass		
Temperature probe tested		Pass		
Anemometer tested		Pass		
Chemical gas sensors calibrated – refer to Calibration Certificate		Pass		
Check main PCB and sensor PCB have the latest software		Yes		
Transducers calibrated		Pass		
As Received Chemical Cell Gas Assessment Performed (CAL16)		Yes		
As Received Gas Assessment Performed (CAL01 or CAL02)		Yes		
Chemical sensor(s) replacement over 2yrs UK and 21mths overseas.		No		
O2 sensor replacement over 3yrs UK, 33mths overseas		Pass		

Instrument Service Report

Unit Type: GA5000	Part Number: GA5KB0B0-301	Date: 29-Jun-2022	Next Service Due: 28-Dec-2022	Customer Name: L K Consult Limited
Serial Number: G507783				
Actions/Investigation Description	Result	Comments		
Verification of Instrument				
Customer Comments				
Returned for full service and calibration				
Accessory Test	Result	Comments		
Charger	Pass	5K Charger Voltage test box indicates charger voltage and current output are in spec.		
Water Trap	Pass	Replaced PTFE Filter, Replaced tubing, Replaced hose barb, Tested ok.		

General Comments/Feedback

Thank you for returning your gas analyser to the QED Service Centre for full service and calibration. We are pleased to inform you that the analyser was received and fully assessed by one of our experienced Service Engineers and no significant faults or issues were observed during the assessment. Our extensive service was carried out, and any necessary components replaced. The analyser has successfully passed all of our rigorous testing and quality checks and has been calibrated using our bespoke, state of the art calibration facility. For further information about how to get the best use from your instrument please visit our YouTube channel <http://www.youtube.com/GeotechTV> and on our Website <https://qedenv.com>

If you require any further assistance with your instrument please email our Technical Support Team at technical@qedenv.co.uk or call us on +44 (0) 333 800 0088 (Monday to Thursday 08.30 - 17.00 & Friday 08.30 - 15.30) UK time zone.

Service Details: Service Scheme	☒	Service Engineer: Ricky Cottrell	Calibration Engineer: Andy Townsend	Approved By: Laura McBride	Signature: 
Standard Service	☐				