



Certificate of Calibration

Customer: Environmental Protection Strategies

Instrument: MiniRAE Lite

Job: Service, Test & Calibration

Serial number: 590-911088

Fleet Number: N/A

Certificate no: 911088/300822

Next calibration due date: 30. Aug 2023

Tested on: 30. Aug 2022

Calibrated for: Isobutylene

<u>Applied Gas Concentration:</u>	<u>Cylinder Reference:</u>	<u>Final Sensor Reading</u>	<u>Accuracy Limits</u>	<u>Pass / Fail</u>
Isobutylene: 100ppm	6593-3-1	100.1ppm	± 10%	Pass

The instrument has been calibrated after re-zeroing and introducing span calibration gas, using gas that is traceable to national standards and has been prepared in accordance with BS EN ISO 6145-6:2017

Calibration Engineer: Chris Shillitoe

Sign:

Quality Assessed by (Print)	Dan	Sign:	
--	-----	--------------	--