Certificate of Calibration

Customer: Tetra Tec

Instrument: MiniRAE LITE

Job: Service, Test & Calibration

Serial number: 595-004573

Fleet Number: -

Certificate no: 004573/260122

Next calibration due date: 26.01.2023

Tested on: 26.01.2022

Calibrated for: Isobutylene

Applied Gas Concentration:	*Cylinder Reference:*	*Initial Sensor Reading*	*Final Sensor Reading*	*Accuracy Limits*
Isobutylene 100ppm	1055/3021	98.7ppm	100.1ppm	±5%

The instrument has been calibrated after re-zeroing and introducing span calibration gas, using gas that is traceable to national standards and has been prepared in accordance with BS EN ISO 6145-6:2017

Calibration Engineer: DANIEL JACQUES [redacted]

Quality Assessed by (Print):	C.S.	**Sign:**	[redacted]